

Alaska Scientific Crime Detection Laboratory

DataMaster Calibration Form

Issued: 4/12/2018
Effective: 4/13/2018

Version: DMCAF 2018 R0
Status: Active

Instrument # 100684

Analyst Derek Walton

Date: 10/04/2019

Reason for Calibration

Internal standard errors

1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Auto
- c. Lot 18370
- d. Expiration Date 12/05/2020

2. Wet Bath Simulator

- a. Serial Number SD198
- b. Date Calibrated 03/06/2019

3. Attach Calibration Factors

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100684

Date: 10/04/2019
Time: 12:49:11

OPERATOR NAME:
DEREK WALTON
CERTIFICATE #: 9508

Ca = 0.1000
CAL = 0.986837 0.800 <= CAL < 1.200
b1 = 0.0009 0.0000 <= b1 < 0.0040
b2 = 0.0048 0.0010 <= b2 < 0.0100
b3 = 0.0009 0.0000 <= b3 < 0.0040
Xq = 0.0929 0.0500 <= Xq < 0.2000
a21 = 1.174117 1.050 <= a21 < 1.300
a31 = 0.531210 0.300 <= a31 < 0.800

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100684

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 10/04/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.52: 0.079
LOT #: AG826101
EXPIRATION: 09/18/2020
TANK PRESSURE: 772 psi

BLANK TEST	0.000	13:13
INTERNAL STANDARD	VERIFIED	13:13
Xq = 0.093	(0.09%)	
EXTERNAL STANDARD	0.077	13:13
X[1] = 0.0773 (0.0006)	(0.0016)	
BLANK TEST	0.000	13:14
EXTERNAL STANDARD	0.077	13:14
X[1] = 0.0768 (0.0002)	(0.0021)	
BLANK TEST	0.000	13:15
EXTERNAL STANDARD	0.077	13:15
X[1] = 0.0772 (0.0002)	(0.0018)	
BLANK TEST	0.000	13:16
EXTERNAL STANDARD	0.077	13:16
X[1] = 0.0772 (0.0009)	(0.0015)	
BLANK TEST	0.000	13:17
EXTERNAL STANDARD	0.077	13:17
X[1] = 0.0771 (0.0007)	(0.0012)	
BLANK TEST	0.000	13:18

Average = 0.0770
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 42.3°C	PASSED
(38.0°C - 50.0°C)	

SETTINGS

Lamp Voltage = 1.85 V	PASSED
(1.11 V - 2.59 V)	
Cooler Voltage = 1.63 V	PASSED
(0.98 V - 2.28 V)	
Bias Voltage = 80 V	PASSED
(48 V - 112 V)	
Chopper Freq = 549 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 29.5 in	

PUMP INFO

Flow Rate = 3.845 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

DETECTOR INFO

PUMP ON	PASSED
(0.001 V <= 0.010 V)	
PUMP OFF	PASSED
(0.002 V <= 0.010 V)	

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD PASSED

Xq = 0.093 (0.00%) (0.00% - 4.00%)