

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.6

Effective: 10/10/2022

Date: 8/2/2023

Instrument #: 100370

Instrument Location: UAF PD

Supervisor Name and Agency: Carlie Glaister SCDL

Describe the Status Change or Issue with the Instrument:

During IFIR it was noticed that repeated tests (5 total) occurred on the same individual on 07/08/2023. The first test resulted in an RFI status message, the following 2 tests resulted in Blank Errors, and then 2 completed tests that resulted in a reported Breath Alcohol Concentration value were performed on the same subject. The Breath Test Supervisor, Andrew Parker was emailed to notify officers that the crime lab does not recommend testing subjects more than 3 times. Additionally, when a breath tests completed successfully and reports a result, it is not recommended to test them a second time. The Breath Test Supervisor was also notified to remind officers to not key their radios in the vicinity of the DataMaster to prevent RFI status messages. After reviewing data during IFIR, the instrument itself appears to be in good working order.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: [Click or tap here to enter text.](#)

For Use by SCDL

Additional Notes

[Click or tap here to enter text.](#)

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

Glaister, Carlie K (DPS)

From: Glaister, Carlie K (DPS)
Sent: Wednesday, August 2, 2023 11:59 AM
To: Andrew Parker
Subject: DataMaster FYI

Hello Andrew, when reviewing status messages I noticed on 7/8 that the same officer performed a breath test on the same subject 5 times in a row. The first 3 resulted in status messages, and then the last 2 were completed breath tests (on the same individual). Generally the crime lab does not recommend testing a subject more than 3 times, and if you get a completed result they should not be tested again. Also remind officers not to key radios in the vicinity of the DataMaster to prevent RFI status messages.

Serial Number	Test Date	Location	Sub	Sut	Su	Ope	Ope	Ope	Subject/Alcohol	Status
100370	6/24/2023 2:09 AM	UAF PD	G...	J...		Jl...	N	C...		RFI
100370	6/24/2023 2:16 AM	UAF PD	G...	J...		Jl...	N	C...		RFI
100370	6/24/2023 2:17 AM	UAF PD	G...	J...		Jl...	N	C...		
100370	6/24/2023 2:23 AM	UAF PD	G...	J...		Jl...	N	C...		
100370	6/30/2023 10:24 PM	UAF PD	V...	F...	L	N...	J	S...	[REDACTED]	Blank Error
100370	6/30/2023 10:30 PM	UAF PD	V...	F...	L	N...	J	S...	0.14200000	
100370	7/1/2023 12:08 AM	UAF PD	D...	H...	N	O...	A	O...		RFI
100370	7/1/2023 12:09 AM	UAF PD	D...	H...	N	O...	A	O...	0.19000000	
100370	7/3/2023 11:04 PM	UAF PD	S...	J...	J	K...	J	C...		
100370	7/8/2023 2:08 AM	UAF PD	N...	D...	B	J...	N	S...		RFI
100370	7/8/2023 2:11 AM	UAF PD	N...	D...	B	J...	N	S...	[REDACTED]	Blank Error
100370	7/8/2023 2:16 AM	UAF PD	N...	D...	B	J...	N	S...	[REDACTED]	Blank Error
100370	7/8/2023 2:22 AM	UAF PD	N...	D...	B	J...	N	S...	0.16000000	
100370	7/8/2023 2:31 AM	UAF PD	N...	D...	B	D...	B	N...	0.16900000	
100370	7/19/2023 2:55 AM	UAF PD	V...	K...	D	J...	N	S...	0.07300000	

Carlie Glaister (Bailey)

Forensic Scientist III – Chemistry Section
State of Alaska Scientific Crime Detection Laboratory
907-269-5689