

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100405

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 08/18/2022

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.38: 0.079
LOT #: AG113402
EXPIRATION: 05/14/2023
TANK PRESSURE: 730 psi

BLANK TEST	0.000	09:17
INTERNAL STANDARD	VERIFIED	09:17
EXTERNAL STANDARD	INTERFERENCE	09:17

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 47.4°C	PASSED

PUMP INFO

Flow Rate = 5.082 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED

File only DM 8/18/22