

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100414

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/19/2024

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.60: 0.079
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 987 psi

BLANK TEST	0.000	15:24
INTERNAL STANDARD	VERIFIED	15:24
EXTERNAL STANDARD	0.078	15:24
BLANK TEST	0.000	15:25
EXTERNAL STANDARD	0.079	15:25
BLANK TEST	0.000	15:26
EXTERNAL STANDARD	0.079	15:26
BLANK TEST	0.000	15:27
EXTERNAL STANDARD	0.079	15:27
BLANK TEST	0.000	15:28
EXTERNAL STANDARD	0.078	15:28
BLANK TEST	0.000	15:29

Average = 0.0786
Std Dev = 0.0005

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 46.3°C	PASSED

PUMP INFO

Flow Rate = 4.910 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File Only JMW 01/03/25