

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMCIF 2015 R1
Status: Active

Instrument S/N 100696

Received From NPAS

Date 2/26/16

Analyst CTF

1. Reason for instrument return:

NPAS Repair - Calibration Failures

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax N/A

4. Upload N/A

5. Update agency information and polling status in DM Host Calling Program N/A

6. Check for VOCs N/A

From NPAS

7. NPAS Repair Paperwork (attached)



PACKING LIST

75013

Special Instructions

Cert Enclosed
 Partial Ship
 Complete Ship

Goods Received in Good Condition

By: _____

Date: _____

SOLD TO: ALASKA DEPT OF PUBLIC SAFETY
 SCIENTIFIC CRIME DETECTION
 4805 DR MARTIN LUTHER KING DR.
 ANCHORAGE AK 99507
 907-269-5582/907-338-6614

SHIP TO: ALASKA DEPT OF PUBLIC SAFETY
 SCIENTIFIC CRIME DETECTION
 4805 DR. MARTIN LUTHER KING DR
 ANCHORAGE AK 99507
 Phone: 907-269-5582 Fax: 907-338-6614

Sales Order	Ship Num	Cust No	Order Date	Tax	Promised	Sales Rep	Customer P.O. Number	Mark Shipment
0050369	0000	AK6056	2/1/2016	E	2/15/2016	HOUSE	COLLEEN	
F.O.B. Point		Ship Date	Shipped Via	Cartons	Weight	Waybill Number	Ins	
DEST		2/18/2016	FED EX				N	

Item	T	QUANTITY			Unit	Part Number/Revision	Description
		Order	B/O	Ship			
001	S	2		2	HR	9DMT-LABOR	REPAIR LABOR
002	S	1		1	EA	37194	MOTOR, 12 X 20.6 MM, 8 MM SHAF T STANDARD SLEEVE BEARING
003	S	1		1	EA	41311	MOTOR ASS'Y, STEPPER
004	S	1		1	EA	37058	CLIP, NYLON CABLE .625 DIA (BT HOLDER)