

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100342

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/16/2023

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.95: 0.077
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 1212 psi

BLANK TEST	0.000	14:17
INTERNAL STANDARD	VERIFIED	14:17
EXTERNAL STANDARD	0.077	14:17

Pump Error

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C	PASSED
Breath Tube = 42.5°C	PASSED

PUMP INFO

Flow Rate = 3.667 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED

File only COB 12/17/23