

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 12-30-2015 Instrument Serial Number: 1000688

Instrument Location: SOLDOTNA AST

Supervisor Name and Agency: TRP. MATTHEW WERTANEN

Describe the Status Change or Issue with the Instrument:

DAMAGED BY PRISONER, RETURNING TO LAB

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 12-30-2015 / 2149

For Use by SCDL

Additional Notes

Instrument removed from service following a report of a prisoner banging his hip against the instrument/desk multiple times and then the DMT having trouble powering up. (see evaluation for email from AST/Soldotna)

Email completed form to info.scdl@toxicology.alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740