

	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Calibration Form - DMCAF2011R1		
	Issuing Authority - Nita Bolz, Scientific Director		
Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active	

DataMaster DMT Calibration Form

Instrument # 100397Date: 5/18/11

## 1. Wet Bath Simulator Solution

- a. Level 0.100  
 b. Manufacturer Guth  
 c. Lot 10260  
 d. Expiration Date 10/25/12  
COB 4/1/10

## 2. Wet Bath Simulator

- a. Serial Number DR-4161  
 b. Date Calibrated 8/23/10

## 3. Attach Calibration Factors

Breath Alcohol Analyst C. O' Bryant

CALIBRATION FACTORS

Alaska Department of Public Safety  
DATAMASTER dmt: 100397

Date: 05/18/2011  
Time: 13:35:39

SUPERVISOR NAME:  
COLLEEN S O'BRYANT  
CERTIFICATE #: 6237

LOT #: GUTH100  
BOTTLE #: 10260  
EXPIRATION: 10/25/2011  
*COB*  
*COB 4/14*

Ca	=	0.1000		
CAL	=	0.983031	0.800	<= CAL < 1.200
b1	=	0.0003	0.0000	<= b1 < 0.0040
b2	=	0.0079	0.0010	<= b2 < 0.0100
b3	=	0.0004	0.0000	<= b3 < 0.0040
Xq	=	0.0932	0.0500	<= Xq < 0.2000
a21	=	1.208582	1.050	<= a21 < 1.300
a31	=	0.521034	0.300	<= a31 < 0.600

*COB*

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100397\B-100397\_ECF-20140

2/19/2014