

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100365

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 05/21/2025

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.27: 0.078
LOT #: AG436502
EXPIRATION: 12/30/2026
TANK PRESSURE: 1177 psi

BLANK TEST	0.000	13:58
INTERNAL STANDARD	VERIFIED	13:58
EXTERNAL STANDARD	0.074	13:58
BLANK TEST	0.000	13:59
EXTERNAL STANDARD	0.074	14:00
BLANK TEST	0.000	14:01
EXTERNAL STANDARD	0.077	14:01
BLANK TEST	0.000	14:02
EXTERNAL STANDARD	INTERFERENCE	14:02

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 45.0°C	PASSED

PUMP INFO

Flow Rate = 4.421 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------

File Only dmt 06/26/25