

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Effective: 10/10/2022

Version: 3.0

Instrument #: 100695

Received From: Whittier PD

Date: 07/18/2023

Analyst: Tatiana Malwoske

Reason for instrument return:

Tank interference on subject tests, RFI on VOC, Diagnostic
Failure on VOC x 3 and elevated chopper wheel values

Change instrument and case location in JusticeTrax Breath Alcohol Database

Ensure change in instrument status form is in JusticeTrax

Upload

Update agency information and polling status in DMHost Calling Program

Check for VOCs

From Intoximeters

Intoximeters Repair Paperwork (attached)