

Alaska Scientific Crime Detection Laboratory

DataMaster Acceptance Form

Issued: 01/27/2012
Effective: 01/27/2012

Version: DMAF2012 R0
Status: Active

Instrument # 100674Date: 10/4/12Analyst: CRF

- Check in instrument in to JusticeTrax Breath Alcohol Database
- Visual Inspection Instrument software would not load. Replaced SE board. See 10/4/12 Evaluation Form - CRF
- Update software Ran USB Fixer
- Print calibration factors (attached) N/A need to calibrate
- NPAS calibration paperwork (attached)
- Barcode Instrument
- State of Alaska Property ID# 1A-F0365

Breath Case B-100674
Sub 100674
QSHS03M3QSDMZ4
DataMaster DMT



Certificate of Calibration

100674
Instrument Serial Number

10260
Standard Reference Number

The above instrument has been calibrated and is hereby certified for use in determining alcohol breath concentration when used in accordance with the operating guidelines by a properly trained operator.

10-27-2011
Date


Technician

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100674\B-100674 ECF-20140

7/10/2014