

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Serial #: 100390

Date: 07/31/2014

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
 TARGET AT 30.16: 0.080
 LOT #: 11313080A1
 EXPIRATION: 06/01/2015

BLANK TEST	0.000	00:09
INTERNAL STANDARD	VERIFIED	00:09
EXTERNAL STANDARD	0.081	00:09
BLANK TEST	0.000	00:10
EXTERNAL STANDARD	0.081	00:10
BLANK TEST	0.000	00:11
EXTERNAL STANDARD	0.081	00:11
BLANK TEST	0.000	00:12
EXTERNAL STANDARD	0.081	00:12
BLANK TEST	0.000	00:13
EXTERNAL STANDARD	0.081	00:13
BLANK TEST	0.000	00:14

Average = 0.0810
 Std Dev = 0.0000

Diagnostic Check

VERSIONS
 DMT: 2.00
 PIC: 3.02
 Modem: 2.3
 Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
 Breath Tube = 45.9°C PASSED

PUMP INFO

Flow Rate = 4.787 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
 PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
 Filter 2 PASSED
 Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz
 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

9/5/2014

Subscribed and sworn before me this 5th day of Sept, 20 14

Nikki Roth
 Nikki Roth, Notary Public
 My Commission Expires With Office



BMB
 8/25/14