

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100382

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 11/15/2020

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.59: 0.079
LOT #: AG923401
EXPIRATION: 08/22/2021
TANK PRESSURE: 881 psi

BLANK TEST	0.000	15:39
INTERNAL STANDARD	VERIFIED	15:39
EXTERNAL STANDARD	0.079	15:39
BLANK TEST	0.000	15:40
EXTERNAL STANDARD	0.079	15:40
BLANK TEST	0.000	15:41
EXTERNAL STANDARD	INTERFERENCE	15:41

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 46.6°C	PASSED

PUMP INFO

Flow Rate = 5.020 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

PASSED

File only. DM 12/15/20