

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100365

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date:09/15/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.53: 0.079
LOT #: AG018102
EXPIRATION: 06/29/2022
TANK PRESSURE: 108 psi

BLANK TEST	0.000	23:28
INTERNAL STANDARD	VERIFIED	23:28
EXTERNAL STANDARD	0.077	23:28
BLANK TEST	0.000	23:29
EXTERNAL STANDARD	0.078	23:29
BLANK TEST	0.000	23:30
EXTERNAL STANDARD	0.077	23:30
BLANK TEST	0.000	23:31

Tank Pressure Too Low

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C	PASSED
Breath Tube = 42.0°C	PASSED

PUMP INFO

Flow Rate = 4.597 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------