

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100699

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/21/2024

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.82: 0.080
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 618 psi

| | | |
|-------------------|--------------|-------|
| BLANK TEST | 0.000 | 02:14 |
| INTERNAL STANDARD | VERIFIED | 02:14 |
| EXTERNAL STANDARD | 0.075 | 02:14 |
| BLANK TEST | 0.000 | 02:15 |
| EXTERNAL STANDARD | 0.075 | 02:15 |
| BLANK TEST | 0.000 | 02:16 |
| EXTERNAL STANDARD | 0.076 | 02:16 |
| BLANK TEST | 0.000 | 02:17 |
| EXTERNAL STANDARD | 0.077 | 02:17 |
| BLANK TEST | 0.000 | 02:18 |
| EXTERNAL STANDARD | INTERFERENCE | 02:18 |

Interference Detected

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

| | |
|-------------------------|--------|
| Sample Chamber = 48.7°C | PASSED |
| Breath Tube = 46.0°C | PASSED |

PUMP INFO

| | |
|-----------------------|--------|
| Flow Rate = 4.301 L/M | PASSED |
|-----------------------|--------|

DETECTOR INFO

| | |
|----------|--------|
| PUMP ON | PASSED |
| PUMP OFF | PASSED |

FILTER INFO

| | |
|----------|--------|
| Filter 1 | PASSED |
| Filter 2 | PASSED |
| Filter 3 | PASSED |

| | |
|-------------------|--------|
| INTERNAL STANDARD | PASSED |
|-------------------|--------|

File only dmt 07/08/24