

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 5/20/2025

Instrument #: Click or tap here to enter text.

100690

Instrument Location: Click or tap here to enter text.

Clear SFS AK ~~9905~~ 99704

Supervisor Name and Agency: Click or tap here to enter text.

Tsgt William C Bailey 268 SFS

Describe the Status Change or Issue with the Instrument:

Click or tap here to enter text.

Back in service

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

20 May 2025 0942

Out of Service Date/Time: Click or tap here to enter text.

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Spoke to William Bailey over the phone. The DataMaster owned by Clear AFB (100255) was returned to replace 100690 (on loan). However, 100255 was displaying detector overflow status messages and was unable to be placed into service. 100690 was returned to service. CKG 5/20/2025

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov).

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor