

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 1/13/2017  
Effective: 1/17/2017

Version: DMCIF 2017 R0  
Status: Active

Instrument S/N 100411

Received From Anchorage PD

Date 12/10/2019

Analyst Derek Walton

1. Reason for instrument return:

Sticky 5 way valve

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in ~~JusticeTrax~~   
on network drive jmw 12/10/19

4. Upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

### From Intoximeters

7. Intoximeters Repair Paperwork (attached)