

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 7/5/2023

Instrument #: 100695

Instrument Location: 660 E Whittier Street

Supervisor Name and Agency: Mark Hager/Whittier Police Department

Describe the Status Change or Issue with the Instrument:

New Instrument

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: Click or tap here to enter text.

Out of Service Date/Time: 6/29/2023 @1100 hours

-----  
**For Use by SCDL**

**Additional Notes**

Click or tap here to enter text.

Instrument was placed out of service and is back at the crime lab due to multiple different issues: tank Interference on Subject test, RFI on a VOC, diagnostic failures on three VOCs, elevated chopper wheel value for all three diagnostic failures on VOCs.  
TLM 07/12/2023

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)  
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor