

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100361

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 12/12/2019

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C PASSED  
Breath Tube = 45.5°C PASSED

### PUMP INFO

Flow Rate = 3.915 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

Detector Overflow

File only DW 12/24/19

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100361

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 12/12/2019

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C PASSED  
Breath Tube = 45.6°C PASSED

### PUMP INFO

Flow Rate = 3.930 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

Detector Overflow

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100361

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 12/12/2019

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C      PASSED  
Breath Tube      = 44.2°C      PASSED

Detector Overflow

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100361

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 12/12/2019

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C PASSED  
Breath Tube = 45.8°C PASSED

### PUMP INFO

Flow Rate = 3.986 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

Detector Overflow

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100361

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 12/14/2019

## External Standard Test Values

## Diagnostic Check

### VERSIONS

DMT: 3.02

PIC: 3.03

Modem: 2.6

Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C

PASSED

Breath Tube = 46.2°C

PASSED

### PUMP INFO

Flow Rate = 3.956 L/M

PASSED

### DETECTOR INFO

PUMP ON

PASSED

PUMP OFF

PASSED

Detector Overflow

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100361

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 12/16/2019

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C PASSED  
Breath Tube = 46.2°C PASSED

### PUMP INFO

Flow Rate = 3.903 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

Detector Overflow

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100361

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 12/16/2019

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 49.0°C	PASSED
Breath Tube = 41.3°C	PASSED

Detector Overflow

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100361

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 12/16/2019

## External Standard Test Values

## Diagnostic Check

VERSIONS  
DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.5°C PASSED  
Breath Tube = 43.3°C PASSED

### PUMP INFO

Flow Rate = 4.152 L/M PASSED

### DETECTOR INFO

PUMP ON PASSED  
PUMP OFF PASSED

Detector Overflow