

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

Version: CIISF 2017 R0
Status: Active

Date: 11-1-21 Instrument Serial Number: 100370

Instrument Location: Ketchikan Police Department, 361 Main Street, Ketchikan, AK 99901

Supervisor Name and Agency: Derek A. McGarrigan, Ketchikan Police Department

Describe the Status Change or Issue with the Instrument:

Datamaster 100370 taken out of service. A replacement instrument was provided by the lab and 100370 is being sent back to the lab.
This Datamaster would not automatically upload VOC's with regularity.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: _____

Out of Service Date/Time: 11-1-21 at 1334

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740