	State of Alaska Scientific Crime Detection Laboratory		
	DataMaster Calibration Form – DMCAF2011R1		
	Issuing Authority – Nita Bolz, Scientific Director		
	Approval date: 03/10/11	Effective date: 03/10/11	Archive date: Active

DataMaster DMT Calibration Form

Instrument # 100399
 Date: 7/26/11


1. Wet Bath Simulator Solution

- a. Level 0.10
- b. Manufacturer Guth
- c. Lot 11050
- d. Expiration Date 4/6/13

2. Wet Bath Simulator

- a. Serial Number DR4177
- b. Date Calibrated 12/16/10

3. Attach Calibration Factors

Breath Alcohol Analyst 
 Charles Foster

1/2 CRT

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100399

Date: 07/26/2011
Time: 16:09:36

SUPERVISOR NAME:
CHARLES FOSTER
CERTIFICATE #: 7734

LOT #: 11050
BOTTLE #: 481
EXPIRATION: 04/06/2013

Ca = 0.1000
CAL = 0.989877 0.800 <= CAL < 1.200
b1 = 0.0003 0.0000 <= b1 < 0.0040
b2 = 0.0055 0.0010 <= b2 < 0.0100
b3 = 0.0002 0.0000 <= b3 < 0.0040
Xq = 0.0842 0.0500 <= Xq < 0.2000
a21 = 1.229529 1.050 <= a21 < 1.300
a31 = 0.416065 0.300 <= a31 < 0.800

2/2 CRT

Final page for report

\

\\ANCSCLSAN01\JTRAX\CASEDOCS\B-100399\B-100399 ECF-20140

2/18/2014