

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 1/13/2017
Effective: 1/17/2017

Version: DMCIF 2017 R0
Status: Active

Instrument S/N 100365

Received From Anchorage PI

Date 10/28/2021

Analyst Carlie Glaister

1. Reason for instrument return:

undefined error on morning of 10/8/21. Later that day status message of 'Test Failed' for 2 tests & breath records did not save to instrument.

2. Change instrument ~~and case location~~ in JusticeTrax Breath Alcohol Database

no case (local)

3. Ensure change in instrument status form is in JusticeTrax

4. Upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

From Intoximeters

7. Intoximeters Repair Paperwork (attached)