VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100691 ***Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*** Date:02/21/2023

External Standard Test Values

EXTERNAL STANDARD INFORMATION
NOMINAL: 0.080
TARGET AT 30.84: 0.082
LOT #: AG113402
EXPIRATION: 05/14/2023
TANK PRESSURE: 499 psi

BLANK TEST	0.000	12:02
INTERNAL STANDARD	VERIFIED	12:02

Gas Flow Error

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C	PASSED
Breath Tube = 48.1°C	PASSED

PUMP INFO

Flow Rate = 5.223 L/M	PASSED

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED

File Only DM 03/24/23