

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMEF2015 R0
Status: Active

Instrument # 100353

Date: 1/12/16

Analyst: Colleen O'Bryant

1. Reason for instrument evaluation

Instrument returned from LAPD because the screen wasn't working.

2. Visual Inspection

3. Troubleshooting/Repair

Attempted to loosen top two screen screws to fix the screen, this did not work.

Checked all connections to the screen - disconnected and reconnected, this did not fix the problem.

Evaluation VOC not possible

4. VOC (attached) N/A

5. Internal Standard value $\leq 3.00\%$ Yes No N/A

6. Verify software version N/A

Return to Factory for Repair

Ready For Calibration

Ready For Certification