

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 6/11/2014
Effective: 6/11/2014

Version: CIISF 2014 R0
Status: Active

Date: 7/27/16 Instrument Serial Number: 100-400

Instrument Location: Palmer PD

Supervisor Name and Agency: Colleen O'Bryant, SCDL

Describe the Status Change or Issue with the Instrument:

gas flow errors on low psi tank. Tank replacement resolved the issue. CBS 7/27/16

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740

O'Bryant, Colleen S (DPS)

From: O'Bryant, Colleen S (DPS)
Sent: Monday, July 11, 2016 1:38 PM
To: Foster, Charles R (DPS); Barnett, Brandi M (DPS)
Subject: RE: Datamaster

Palmer only has 199psi left in their tank, that could be the issue. I'll take a new one out tonight.

**Colleen O'Bryant – Forensic Scientist III
Blood and Breath Alcohol**

**907-269-5783 (direct)
907-350-2111 (cell)**

From: DPS SCDL Toxicology (DPS sponsored)
Sent: Monday, July 11, 2016 11:12 AM
To: Foster, Charles R (DPS); O'Bryant, Colleen S (DPS); Barnett, Brandi M (DPS)
Subject: FW: Datamaster

From: Lucas N. Szipszky [lszipszky@PalmerPolice.com]
Sent: Monday, July 11, 2016 9:18 AM
To: DPS SCDL Toxicology (DPS sponsored)
Subject: FW: Datamaster

My bad, datamaster number 100400

**Sergeant Luke Szipszky
Palmer Police Department
Phone: (907)745-4811
Fax: (907)746-2314**

From: Lucas N. Szipszky
Sent: Monday, July 11, 2016 9:23 AM
To: 'dps.scdl.toxicology@alaska.gov' <dps.scdl.toxicology@alaska.gov>
Subject: Datamaster

Hello,
We've been experiencing a high number of Gas Flow Errors. It happens on the first time they try and start a subject test, but clears after a "redo." Is there anything we can do on our end or is this due to the new bottles?

**Sergeant Luke Szipszky
Palmer Police Department
Phone: (907)745-4811
Fax: (907)746-2314**