

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100670

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 07/14/2016

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 30.11: 0.081
LOT #: AG608504
EXPIRATION: 03/25/2018
TANK PRESSURE: 966 psi

BLANK TEST	0.000	08:43
INTERNAL STANDARD	VERIFIED	08:43
EXTERNAL STANDARD	0.081	08:44
BLANK TEST	0.000	08:44
EXTERNAL STANDARD	0.081	08:45
BLANK TEST	0.000	08:46
EXTERNAL STANDARD	0.082	08:46
BLANK TEST	0.000	08:47
EXTERNAL STANDARD	0.081	08:47
BLANK TEST	0.000	08:48
EXTERNAL STANDARD	0.081	08:48
BLANK TEST	0.000	08:49

Average = 0.0812
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 47.9°C PASSED

PUMP INFO

Flow Rate = 4.882 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Charles R. Foster, after being first duly sworn, depose and state as follows:

(1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.

(2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.

(3) I am the Scientific Director of the State Breath Alcohol Program.

(4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.


(5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.

(6) The referenced instrument is certified for evidentiary use in the State of Alaska.


 8/5/16

Charles R. Foster
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 5th day of Aug. 20 16



Nikki Roth, Notary Public
My Commission Expires With Office




7/15/16