

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100685

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/15/2019

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.85: 0.080
LOT #: AG735001
EXPIRATION: 12/16/2019
TANK PRESSURE: 656 psi

BLANK TEST	0.000	10:30
INTERNAL STANDARD	VERIFIED	10:30
EXTERNAL STANDARD	0.079	10:30
BLANK TEST	0.000	10:31
EXTERNAL STANDARD	0.080	10:32
BLANK TEST	0.000	10:32
EXTERNAL STANDARD	0.080	10:33
BLANK TEST	0.000	10:34
EXTERNAL STANDARD	0.080	10:34
BLANK TEST	0.000	10:35
EXTERNAL STANDARD	0.080	10:35
BLANK TEST	0.000	10:36

Average = 0.0798
Std Dev = 0.0004

Diagnostic Check

VERSIONS
DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C PASSED
Breath Tube = 44.2°C PASSED

PUMP INFO

Flow Rate = 5.636 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD

PASSED

File only. Instrument taken out
of service on 02/15/2019 for multiple
status messages. DMW 03/15/2019