

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100353

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 11/04/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 27.82: 0.074
LOT #: AG113402
EXPIRATION: 05/14/2023
TANK PRESSURE: 1232 psi

BLANK TEST	0.000	12:45
INTERNAL STANDARD	VERIFIED	12:45
EXTERNAL STANDARD	0.071	12:45
BLANK TEST	0.000	12:46
EXTERNAL STANDARD	0.073	12:46
BLANK TEST	0.000	12:47
EXTERNAL STANDARD	0.074	12:47
BLANK TEST	0.000	12:48
EXTERNAL STANDARD	0.073	12:48
BLANK TEST	0.000	12:49
EXTERNAL STANDARD	0.073	12:49
BLANK TEST	0.000	12:50

Average = 0.0728
Std Dev = 0.0011

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.07
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.1°C PASSED
Breath Tube = 48.0°C PASSED

PUMP INFO

Flow Rate = 5.083 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

File Only: Instrument removed from service ^{BMB} 12/9/21