

Alaska Scientific Crime Detection Laboratory  
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 10/7/2022

Instrument #: 100688

Instrument Location: Sitka, AK

Supervisor Name and Agency: Nathan Duce, Alaska State Troopers

Describe the Status Change or Issue with the Instrument:

Datamaster DMT wouldn't run and when it was rebooted it displayed the following status messages; Filter wheel error. After going into tech mode the DMT would fail randomly on each of the 3 filters. After returning to the main screen and attempting to reuse the DMT would display this status message; Internal Standard Error.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: [Click or tap here to enter text.](#)

Out of Service Date/Time: 10/7/2022 no longer using for training

-----  
**For Use by SCDL**

Additional Notes

[Click or tap here to enter text.](#)

**Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)**  
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor