

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100695

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 12/07/2018

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 28.98: 0.077
LOT #: AG735001
EXPIRATION: 12/16/2019
TANK PRESSURE: 1168 psi

BLANK TEST	0.000	08:45
INTERNAL STANDARD	VERIFIED	08:45
EXTERNAL STANDARD	0.077	08:45
BLANK TEST	0.000	08:46
EXTERNAL STANDARD	0.078	08:46
BLANK TEST	0.000	08:47
EXTERNAL STANDARD	0.078	08:47
BLANK TEST	0.000	08:48
EXTERNAL STANDARD	0.078	08:48
BLANK TEST	0.000	08:49
EXTERNAL STANDARD	0.078	08:50
BLANK TEST	0.000	08:50

Average = 0.0778
Std Dev = 0.0004

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.7°C PASSED
Breath Tube = 47.5°C PASSED

PUMP INFO

Flow Rate = 4.244 L/M PASSED

DETECTOR INFO

PUMP ON PASSED
PUMP OFF PASSED


FILTER INFO

Filter 1 PASSED
Filter 2 PASSED
Filter 3 PASSED

INTERNAL STANDARD PASSED

I, Charles R. Foster, after being first duly sworn, depose and state as follows:


- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.




Charles R. Foster
Scientific Director
State Breath Alcohol Program


1/25/19

Subscribed and sworn before me this 25th day of Jan. 20 19

Kurt R. Palmer Sr., Notary Public
My Commission Expires With Office


1/25/19




Nikki Roth, Notary Public
My Commission Expires With Office

DW 12/19/18