

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Issued: 12/15/2017
Effective: 12/15/2017

8107 8 2 AM

Version: CIISF 2017 R0
Status: Active

Date: 05-25-2018 Instrument Serial Number: 100681

Instrument Location: Whittier

Supervisor Name and Agency: Mark Hager Whittier Police Dept.

Describe the Status Change or Issue with the Instrument:

Not connecting to phone line + USB Not working

Instrument was working had wrong USB Thumb Drive + Phone line

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 1807 05-25-2018

Out of Service Date/Time: 0700 05-25-2018

For Use by SCDL

Additional Notes

Instrument unable to upload via phone line or USB. Mark Hager brought instrument to lab. Able to upload via phone line at lab and USB depending on USB used; issue is not with instrument. Mark Hager returned instrument to Whittier. JMH 05/30/18

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740