

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 4.0

Effective: 8/10/2022

Date: 3/28/2023

Instrument #: 100424

Instrument Location: Eielson AFB

Supervisor Name and Agency: Derek Walton - SCDL

Describe the Status Change or Issue with the Instrument:

I spoke to Eric Hampton (operator) on the phone. He indicated that the instrument had Ambient Fails messages "a couple weeks ago" and that at least one breath test subject was taken to Ft. Wainwright to use the DataMaster located there. I notified him that the last VOC we have from the Eielson instrument (as of 3/28/23) is 01/03/23 and that the instrument cannot communicate via modem line for an unknown reason. Therefore, I could not look at this test via DMHost. He ran a full non-drinking subject test on 3/28/23 and the instrument performed normally. We discussed room ventilation for their particular location and he agreed to remind officers to ventilate the room as needed and only bring the subject to the instrument when ready to blow.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to dps.scdl.toxicology@alaska.gov
For questions contact the Breath Alcohol Program at 907-269-5740

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor