

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100379

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 01/19/2025

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080

TARGET AT 29.56: 0.079

LOT #: AG335202

EXPIRATION: 12/18/2025

TANK PRESSURE: 1216 psi

BLANK TEST 0.000 17:19

INTERNAL STANDARD VERIFIED 17:19

EXTERNAL STANDARD 0.071 17:20

Standard Out of Range

Diagnostic Check

VERSIONS

DMT: 3.02

PIC: 3.02

Modem: 2.6

Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C PASSED

Breath Tube = 42.6°C PASSED

PUMP INFO

Flow Rate = 4.422 L/M PASSED

DETECTOR INFO

PUMP ON PASSED

PUMP OFF PASSED

FILTER INFO

Filter 1 PASSED

Filter 2 PASSED

Filter 3 PASSED

INTERNAL STANDARD PASSED

File only CRF 2/5/25