

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument  
State of Alaska*

Serial #: 100395

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 05/18/2021

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.53: 0.079  
LOT #: AG923401  
EXPIRATION: 08/22/2021  
TANK PRESSURE: 788 psi

BLANK TEST	0.000	17:54
INTERNAL STANDARD	VERIFIED	17:54
EXTERNAL STANDARD	INTERFERENCE	17:55

Interference Detected

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.8°C	PASSED
Breath Tube = 48.1°C	PASSED

### PUMP INFO

Flow Rate = 4.483 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

PASSED

File only JM 06/15/21