

# Alaska Scientific Crime Detection Laboratory

## DataMaster Evaluation Form

Issued: 7/10/2015  
Effective: 7/10/2015

Version: DMEF2015 R0  
Status: Active

Instrument # 100413

Date: 8/24/15

Analyst: Colleen O'Bryant

1. Reason for instrument evaluation

Instrument originally returned for internal standard errors and also experienced fatal errors while trying to access the records menu.

2. Visual Inspection

3. Troubleshooting/Repair

No longer experiencing fatal errors after reinstalling 3.00 and then 3.01 software.

Other instruments had experienced similar issues when software was updated from 2.00 to 3.01 without 3.00 being uninstalled between them. Due to success with other instruments I pulled this instrument back from being returned to NPAS to attempt the software update + it worked.

4. VOC (attached)

5. Internal Standard value  $\leq 3.00\%$  Yes  No

6. Verify software version

Return to Factory for Repair

Ready For Calibration

Ready For Certification

# VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100413

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 08/24/2015

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 29.59: 0.079  
LOT #: 09214080A2  
EXPIRATION: 05/01/2016  
TANK PRESSURE: 340 psi

BLANK TEST	0.000	08:42
INTERNAL STANDARD	VERIFIED	08:42
Xq = 0.093 (2.99%)		
EXTERNAL STANDARD	0.077	08:43
X[1] = 0.0767 (-0.0003) (0.0010)		
BLANK TEST	0.000	08:44
EXTERNAL STANDARD	0.077	08:44
X[1] = 0.0768 (-0.0009) (0.0012)		
BLANK TEST	0.000	08:45
EXTERNAL STANDARD	0.077	08:45
X[1] = 0.0767 (-0.0003) (0.0000)		
BLANK TEST	0.000	08:46
EXTERNAL STANDARD	0.077	08:46
X[1] = 0.0769 (-0.0003) (0.0006)		
BLANK TEST	0.000	08:47
EXTERNAL STANDARD	0.077	08:47
X[1] = 0.0769 (-0.0009) (-0.0003)		
BLANK TEST	0.000	08:48

Average = 0.0770  
Std Dev = 0.0000

## Diagnostic Check

### VERSIONS

DMT: 3.01  
PIC: 3.02  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.7°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 45.2°C	PASSED
(39.0°C - 50.0°C)	

### SETTINGS

Lamp Voltage = 1.80 V	PASSED
(1.09 V - 2.55 V)	
Cooler Voltage = 1.59 V	PASSED
(0.98 V - 2.30 V)	
Bias Voltage = 79 V	PASSED
(47 V - 111 V)	
Chopper Freq = 538 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 29.6 in	

### PUMP INFO

Flow Rate = 5.198 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

### DETECTOR INFO

PUMP ON	PASSED
(0.002 V <= 0.010 V)	
PUMP OFF	PASSED
(0.002 V <= 0.010 V)	

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

Xq = 0.093	3.19%	(0.00% - 4.00%)	PASSED
------------	-------	-----------------	--------

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

Nita J. Bolz  
Scientific Director  
State Breath Alcohol Program

Subscribed and sworn before me this \_\_\_\_\_ day of \_\_\_\_\_, 20 \_\_\_\_\_