

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 12/15/15 Instrument Serial Number: 100384

Instrument Location: Juneau Police Dept

Supervisor Name and Agency: Sterling Sulistrom JPD

Describe the Status Change or Issue with the Instrument:

Received from LAB

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 12/15/15 12:00 pm

Out of Service Date/Time: \_\_\_\_\_

For Use by SCDL

Additional Notes

\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_  
\_\_\_\_\_

Email completed form to \_\_\_\_\_

For questions contact the Breath Alcohol Program at 907-269-5740