

Alaska Scientific Crime Detection Laboratory

DataMaster Evaluation Form

Issued: 7/10/2015
Effective: 7/10/2015

Version: DMEF2015 R0
Status: Active

Instrument # 100353

Date: 7/26/16

Analyst: CWF

1. Reason for instrument evaluation

Returned from NPAS - Screen wasn't working

2. Visual Inspection

3. Troubleshooting/Repair

Screen works properly now.

4. VOC (attached)

5. Internal Standard value $\leq 3.00\%$ Yes No

6. Verify software version

Return to Factory for Repair

Ready For Calibration

Ready For Certification

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100353

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 02/26/2016

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.00: 0.078
LOT #: 01315080A3
EXPIRATION: 02/05/2017
TANK PRESSURE: 637 psi

BLANK TEST	0.000	16:03
INTERNAL STANDARD	VERIFIED	16:03
Xq = 0.093 (1.49%)		
EXTERNAL STANDARD	0.077	16:03
X[1] = 0.0774 (-0.0023) (0.0002)		
BLANK TEST	0.000	16:04
EXTERNAL STANDARD	0.077	16:04
X[1] = 0.0773 (-0.0016) (0.0009)		
BLANK TEST	0.000	16:05
EXTERNAL STANDARD	0.077	16:06
X[1] = 0.0774 (-0.0026) (0.0005)		
BLANK TEST	0.000	16:06
EXTERNAL STANDARD	0.077	16:07
X[1] = 0.0772 (-0.0017) (0.0001)		
BLANK TEST	0.000	16:08
EXTERNAL STANDARD	0.077	16:08
X[1] = 0.0773 (-0.0017) (0.0000)		
BLANK TEST	0.000	16:09

Average = 0.0770
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 49.0°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 47.5°C	PASSED
(38.0°C - 50.0°C)	

SETTINGS

Lamp Voltage = 2.17 V	PASSED
(1.30 V - 3.04 V)	
Cooler Voltage = 2.10 V	PASSED
(1.26 V - 2.94 V)	
Bias Voltage = 80 V	PASSED
(48 V - 112 V)	
Chopper Freq = 526 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 29.0 in	

PUMP INFO

Flow Rate = 4.810 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

DETECTOR INFO

PUMP ON	PASSED
(0.002 V <= 0.010 V)	
PUMP OFF	PASSED
(0.002 V <= 0.010 V)	

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

Xq = 0.094	0.56%	(0.00% - 4.00%)	PASSED
------------	-------	-----------------	--------