

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 1/13/2017  
Effective: 1/17/2017

Version: DMCIF 2017 R0  
Status: Active

Instrument S/N 100699

Received From APD

Date 02/14/2020

Analyst Derek Walter

1. Reason for instrument return:

Gas flow errors

2. Change instrument ~~and case~~ location in JusticeTrax Breath Alcohol Database

*Not returned with case Jm 02/11/20*

3. Ensure change in instrument status form is in JusticeTrax

4. Upload

5. Update agency information and polling status in DM Host Calling Program

6. Check for VOCs

### From Intoximeters

7. Intoximeters Repair Paperwork (attached)