

Alaska Scientific Crime Detection Laboratory

DataMaster Check In Form

Issued: 2/2/2015
Effective: 2/2/2015

Version: DMCIF 2015 R0
Status: Active

Instrument S/N 100348

Received From Sand Point PD

Date 6/9/15

Analyst Colleen O'Bryant

1. Reason for instrument return:

Detector overflow on April VOC.

2. Change instrument and case location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Visual Inspection

5. Upload

6. Update agency information and polling status in DM Host Calling Program

7. Check for VOCs

From NPAS

8. NPAS Repair Paperwork (attached)

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100348

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 04/16/2015

External Standard Test Values

Diagnostic Check

```

VERSIONS
DMT: 3.00
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C          PASSED
(44.0°C - 52.0°C)
Breath Tube    = 47.8°C          PASSED
(39.0°C - 50.0°C)

SETTINGS
Lamp Voltage   = 2.06 V          PASSED
(1.25 V - 2.93 V)
Cooler Voltage = 1.82 V          PASSED
(1.01 V - 2.37 V)
Bias Voltage   = 79 V           PASSED
(47 V - 109 V)
Chopper Freq   = 546 Hz         PASSED
(475 Hz - 575 Hz)
Barometer      = 29.5 in

PUMP INFO
Flow Rate     = 4.375 L/M        PASSED
(3.500 L/M - 6.500 L/M)

DETECTOR INFO
PUMP ON       = 0.003 V <= 0.010 V PASSED
PUMP OFF      = 0.003 V <= 0.010 V PASSED
    
```

Detector Overflow

I, Nita J. Bolz, after being first duly sworn, depose and state as follows:

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the Breath Alcohol Program, which includes assuring that instruments are calibrated and maintaining program records.
- (5) The above is a true and accurate verification of calibration, which is performed by the instrument's software, as specified by the State Breath Alcohol Program. Verification of calibration is a regularly conducted and regularly recorded activity of the State Breath Alcohol Program.
- (6) The referenced instrument is certified for evidentiary use in the State of Alaska.

 Nita J. Bolz
 Scientific Director
 State Breath Alcohol Program

Subscribed and sworn before me this ____ day of _____, 20 ____

Technician Screen

DATAMASTER dmt: 100348

Date: 06/09/2015
Time: 8:58 AM

Temperatures (°C)
Sample Cell: 48.7°C
Breath Tube: 44.9°C

Current Barometer 29.3 in
Volume (Ltr) 0.00

Settings
Lamp: 2.06 V
Bias: 80 V
Cooler: 1.82 V
Chopper: 530 Hz

Voltages (V)

Flow: 1.03 V
Detector: 1.973 V

*As found detector
voltage after returning
from the field.*

COB 6/9/15