

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 1/14/26

Instrument #: 100681

Instrument Location: SCDL

Supervisor Name and Agency: C. O Bryant

Describe the Status Change or Issue with the Instrument:

*Instrument sent to Cordova PD to replace their instrument that needed preventative maintenance. When trying to set it up the officer heard a pop & the screen went black + wouldn't turn on.
Instrument was not put in service + returned.*

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: NA

Out of Service Date/Time: NA

For Use by SCDL

Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

**[Draft] Fw: DATA MASTER****From**

Draft saved Wed 1/14/2026 3:38 PM

From: Blake Reece <BReece@cordovapdak.gov>
Sent: Sunday, January 11, 2026 3:56 PM
To: Barnett, Brandi M (DPS) <brandi.barnett@alaska.gov>
Subject: DATA MASTER

You don't often get email from breece@cordovapdak.gov. [Learn why this is important](#)

CAUTION: This email originated from outside the State of Alaska mail system. Do not click links or open attachments unless you recognize the sender and know the content is safe.

Good morning!

We recently received a new datamaster to replace our old datamaster that needs to be returned. This is just a routine return, and the old data master needs to go back for routine maintenance. The new data master we received is not functioning. The device was plugged in, went through a standard boot up process, and then the machine clicked and no longer has power. We attempted to boot the device using different outlets and power cables to get the machine to boot up but this did not work. I have set up the old data master device, and I have the new one (that is not functioning) put back into the hard case. I have a dispatcher who is attending a training located at the crime lab this week. I could send the broken data master unit up with him if you guys are available to receive it. If needed, you are welcome to call me at 907-831-9052. Thank you for you're time!

Blake Reece
Cordova Police Department
Patrol Officer
breece@cordovakapd.gov
Dispatch 907-424-6100 | Fax 907-424-6125

