

Alaska Scientific Crime Detection Laboratory

Change in Instrument Status Form

Effective: 11/10/2021

Version: 2.0

Date: 1/10/22 Instrument #: 100421

Instrument Location: JBFR

Supervisor Name and Agency: C. O. Bryant / SCDL

Describe the Status Change or Issue with the Instrument:

The instrument was brought to the lab to troubleshoot uploading. JBFR, Breath Test Supervisor, Scott Gardner, had also uploaded files from the instrument onto a USB drive & was unable to remove them from the drive. I pulled files from 12.12.2021 - 01.10.2022 from the instrument & was able to transfer them to DMHost at the lab. I also performed a non-drinking subject test & diagnostic tests (two each) & was able to transfer files to a USB drive & retrieve them as well as upload the DMT via modem line.
COB 1/10/22

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 1/10/22

Out of Service Date/Time: _____

For Use by SCDL	Additional Notes

Email completed form to dps.scdl.toxicology@alaska.gov

For questions contact the Breath Alcohol Program at 907-269-5740