



7 Eastgate Dr. • P.O. Box 790 • Jacksonville, IL 62651-0790  
217-245-2183 • Fax: 217-243-7634 • www.ilmoproducts.com


## Certificate of Analysis

**Certificate ID:** 1459  
**Part #:** BAC105L080T  
**Cylinder Size:** 105L  
**Lot Number:** 07511080A1  
**Expiration:** 3/1/2013

### 0.080 BAC (For use with breath alcohol testing instruments)

**Contents:** 105 Liters @ 1000 psig 70°F (21°C)

Component:	Concentration:	Accuracy:	Method:
Ethanol	208.4 ppm	+/- 0.002 or 2%	NDIR
Nitrogen	Balance	BAC whichever is greater	

  
Specialty Gas Lab Manager

3-24-2011  
Date

Distributed by: CMI Inc.  
316 East Ninth Street  
Owensboro, KY 42303  
Phone 866-835-0690  
[www.alcoholtest.com](http://www.alcoholtest.com)



SUPERVISOR TEST

Alaska Department of Public Safety  
 DATAMASTER dmt: 100359

Date: 04/12/2011  
 Time: 08:57:38

SUPERVISOR NAME:  
 COLLEEN S OBRYANT  
 OPERATOR'S #: 07511080A1

EXTERNAL STANDARD INFORMATION  
 NOMINAL: 0.080  
 TARGET AT 29.59: 0.079

BLANK TEST	0.000	08:59
INTERNAL STANDARD	VERIFIED	09:00
EXTERNAL STANDARD	0.080	09:00
BLANK TEST	0.000	09:01
EXTERNAL STANDARD	0.080	09:01
BLANK TEST	0.000	09:02
EXTERNAL STANDARD	0.080	09:02
BLANK TEST	0.000	09:03
EXTERNAL STANDARD	0.081	09:03
BLANK TEST	0.000	09:04
EXTERNAL STANDARD	0.080	09:04
BLANK TEST	0.000	09:05

Average = 0.0802  
 Std Dev = 0.0004

*ILMO 07511080A1*

*Level 0.080*

*Exp 3/1/2013*

