

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100400

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date:09/23/2024

External Standard Test Values

Diagnostic Check

```
VERSIONS
DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C      PASSED
Breath Tube    = 48.1°C      PASSED

PUMP INFO
Flow Rate = 4.850 L/M      PASSED

DETECTOR INFO
PUMP ON          PASSED
PUMP OFF        PASSED

FILTER INFO
Filter 1        PASSED
Filter 2        PASSED
Filter 3        PASSED

INTERNAL STANDARD      FAILED

Diagnostic Check Failed
```

File only dm 10/11/24