

Alaska Scientific Crime Detection Laboratory
Change in Instrument Status Form

Version: 6.0

Effective: 1/17/2024

Date: 9/27/2024

Instrument #: 100368

Instrument Location: Skagway Police Department

Supervisor Name and Agency: James Michels, Skagway Police

Describe the Status Change or Issue with the Instrument:

In Service, replaced DataMaster 100365

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 9/27/2024 9:55 am

Out of Service Date/Time: Click or tap here to enter text.

For Use by SCDL

Additional Notes

Click or tap here to enter text.

Email completed form to dps.scdl.toxicology@alaska.gov.

Page 1 of 1

All printed copies are uncontrolled.

Approved by: Chemistry Supervisor

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100368

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date:09/27/2024

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.34: 0.078
LOT #: AG310901
EXPIRATION: 04/19/2025
TANK PRESSURE: 690 psi

BLANK TEST	0.000	09:46
INTERNAL STANDARD	VERIFIED	09:46
EXTERNAL STANDARD	0.077	09:46
BLANK TEST	0.000	09:47
EXTERNAL STANDARD	0.077	09:47
BLANK TEST	0.000	09:48
EXTERNAL STANDARD	0.077	09:48
BLANK TEST	0.000	09:49
EXTERNAL STANDARD	0.077	09:49
BLANK TEST	0.000	09:50
EXTERNAL STANDARD	0.077	09:50
BLANK TEST	0.000	09:51

Average = 0.0770
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.03
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.6°C	PASSED
Breath Tube = 39.5°C	PASSED

PUMP INFO

Flow Rate = 4.444 L/M	PASSED
-----------------------	--------

DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD	PASSED
-------------------	--------