

# VERIFICATION OF CALIBRATION REPORT

*of DataMaster dmt Breath Test Instrument*

*State of Alaska*

Serial #: 100686

*Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program*

Date: 10/24/2020

## External Standard Test Values

### EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080  
TARGET AT 30.45: 0.081  
LOT #: AG018102  
EXPIRATION: 06/29/2022  
TANK PRESSURE: 1202 psi

BLANK TEST	0.000	12:02
INTERNAL STANDARD	VERIFIED	12:02
EXTERNAL STANDARD	0.078	12:02
BLANK TEST	0.000	12:03

RFI DETECTED

## Diagnostic Check

### VERSIONS

DMT: 3.02  
PIC: 3.03  
Modem: 2.6  
Questions: 2.2

### TEMPERATURES

Sample Chamber = 48.9°C	PASSED
Breath Tube = 46.5°C	PASSED

### PUMP INFO

Flow Rate = 4.790 L/M	PASSED
-----------------------	--------

### DETECTOR INFO

PUMP ON	PASSED
PUMP OFF	PASSED

### FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

### INTERNAL STANDARD

PASSED