

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 8-2-17 Instrument Serial Number: 100697

Instrument Location: ~~University~~ University Police Dept Anchorage

Supervisor Name and Agency: Kevin Stevenson

Describe the Status Change or Issue with the Instrument:

having Problem with Dial up (modem)

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 8-2-17 12:00

Out of Service Date/Time: 8-2-17 10:05 AM

For Use by SCDL

Additional Notes

Troubleshooting the functionality of the instrument  
modem versus agency modem line. Successfully  
uploaded the instrument @ SCDL and  
returned the instrument to service. Instructed  
KS to get in contact with UAA IT (or equivalent)  
to troubleshoot UAA modem line. CRB 8/2/17

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740