

# Alaska Scientific Crime Detection Laboratory

## Change in Instrument Status Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: CIISF 2014 R0  
Status: Active

Date: 11.13.15 Instrument Serial Number: 100691

Instrument Location: VALDEZ POLICE DEPARTMENT

Supervisor Name and Agency: Ofc. CHAD CLEMENTS VPD

Describe the Status Change or Issue with the Instrument:

- REPLACEMENT DATAMASTER -

OLD DATAMASTER NEEDED AT CRIME LAB FOR PM

When I turned on new Datamaster it froze up several times before it finally performed VOC test.

If instrument was placed in service or removed from service, fill out the appropriate sections below.

In Service Date/Time: 11.13.15 / 0837

Out of Service Date/Time: \_\_\_\_\_

For Use by SCDL

Additional Notes

Instrument was freezing when put in service. It had not been plugged in for long so I instructed officer Clements to let it sit a bit longer and try again. There were no more issues.

BMB 11/20/15

Email completed form to [dps.scdl.toxicology@alaska.gov](mailto:dps.scdl.toxicology@alaska.gov)

For questions contact the Breath Alcohol Program at 907-269-5740