

# Alaska Scientific Crime Detection Laboratory

## DataMaster Check In Form

Issued: 6/11/2014  
Effective: 6/11/2014

Version: DMCIF2014 R0  
Status: Active

Instrument S/N 100674

Received From Kodiak PD

Date 1/20/15

Analyst Brandi Barnett

1. Reason for instrument return:  
Broken 5-way valve. Won't accept sample

2. Change instrument location in JusticeTrax Breath Alcohol Database

3. Ensure change in instrument status form is in JusticeTrax

4. Visual Inspection

5. Upload

6. Update agency information and polling status in DM Host Calling Program

7. Check for VOCs

From NPAS

8. NPAS Repair Paperwork (attached)