

Alaska Scientific Crime Detection Laboratory

DataMaster Calibration Form

Issued: 4/12/2018
Effective: 4/13/2018

Version: DMCAF 2018 R0
Status: Active

Instrument # 100687

Analyst Derek Walton

Date: 03/08/21

Reason for Calibration

Filter 2 readings close to maximum on VDC

1. Wet Bath Simulator Solution

- a. Level 0.100
- b. Manufacturer Lifeloc
- c. Lot 20AP41118
- d. Expiration Date 01/15/22

2. Wet Bath Simulator

- a. Serial Number SD1199
- b. Date Calibrated 05/07/20

3. Attach Calibration Factors

CALIBRATION FACTORS

Alaska Department of Public Safety
DATAMASTER dmt: 100687

Date: 03/08/2021
Time: 14:58:17

OPERATOR NAME:
DEREK WALTON
CERTIFICATE #: 9508

Ca = 0.1000
CAL = 0.971467 0.800 <= CAL < 1.200
b1 = 0.0004 0.0000 <= b1 < 0.0040
b2 = 0.0052 0.0010 <= b2 < 0.0100
b3 = 0.0005 0.0000 <= b3 < 0.0040
Xq = 0.0917 0.0500 <= Xq < 0.2000
a21 = 1.152912 1.050 <= a21 < 1.300
a31 = 0.529376 0.300 <= a31 < 0.800

VERIFICATION OF CALIBRATION REPORT

of DataMaster dmt Breath Test Instrument

State of Alaska

Serial #: 100687

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

Date: 03/08/2021

External Standard Test Values

EXTERNAL STANDARD INFORMATION

NOMINAL: 0.080
TARGET AT 29.92: 0.080
LOT #: AG923401
EXPIRATION: 08/22/2021
TANK PRESSURE: 829 psi

BLANK TEST	0.000	15:22
INTERNAL STANDARD	VERIFIED	15:22
Xq = 0.092 (0.76%)		
EXTERNAL STANDARD	0.077	15:22
X[1] = 0.0774 (-0.0004) (0.0023)		
BLANK TEST	0.000	15:23
EXTERNAL STANDARD	0.077	15:24
X[1] = 0.0770 (-0.0013) (0.0020)		
BLANK TEST	0.000	15:24
EXTERNAL STANDARD	0.077	15:25
X[1] = 0.0770 (-0.0013) (0.0019)		
BLANK TEST	0.000	15:26
EXTERNAL STANDARD	0.077	15:26
X[1] = 0.0768 (-0.0016) (0.0019)		
BLANK TEST	0.000	15:27
EXTERNAL STANDARD	0.077	15:27
X[1] = 0.0766 (-0.0016) (0.0018)		
BLANK TEST	0.000	15:28

Average = 0.0770
Std Dev = 0.0000

Diagnostic Check

VERSIONS

DMT: 3.02
PIC: 3.02
Modem: 2.6
Questions: 2.2

TEMPERATURES

Sample Chamber = 48.8°C	PASSED
(44.0°C - 52.0°C)	
Breath Tube = 45.3°C	PASSED
(38.0°C - 50.0°C)	

SETTINGS

Lamp Voltage = 1.59 V	PASSED
(0.95 V - 2.23 V)	
Cooler Voltage = 1.61 V	PASSED
(0.97 V - 2.25 V)	
Bias Voltage = 80 V	PASSED
(48 V - 112 V)	
Chopper Freq = 535 Hz	PASSED
(475 Hz - 575 Hz)	
Barometer = 29.9 in	

PUMP INFO

Flow Rate = 5.026 L/M	PASSED
(3.500 L/M - 6.500 L/M)	

DETECTOR INFO

PUMP ON	PASSED
(0.002 V <= 0.010 V)	
PUMP OFF	PASSED
(0.001 V <= 0.010 V)	

FILTER INFO

Filter 1	PASSED
Filter 2	PASSED
Filter 3	PASSED

INTERNAL STANDARD

Xq = 0.092 0.10%	PASSED
(0.00% - 4.00%)	