

CALIBRATION FACTORS

-----  
Alaska Department of Public Safety  
DATAMASTER dmt: 100667  
-----

Date: 02/10/2020  
Time: 09:44:12

OPERATOR NAME:  
INTOX I METERS  
CERTIFICATE #: 1

Ca = 0.2000  
CAL = 0.992420    0.800 <= CAL < 1.200  
b1 = 0.0008    0.0000 <= b1 < 0.0040  
b2 = 0.0075    0.0010 <= b2 < 0.0100  
b3 = 0.0008    0.0000 <= b3 < 0.0040  
Xq = 0.1056    0.0500 <= Xq < 0.2000  
a21 = 1.215045    1.050 <= a21 < 1.300  
a31 = 0.473071    0.300 <= a31 < 0.800